

1395 Sandhill Drive, Ancaster, Ontario, Canada L9G 4V5 Local: (905)304·9896 Toll Free: (800)903·6020 Fax: (905)304·9897 Email: info@affinitybiologicals.com

Certificate of Analysis for Fibrinogen (Fg) Deficient Human Plasma

Product:

Fg Deficient Plasma

Lot #: DP1-0026

Expiry: 08/2014

Preparation:

Pooled normal citrated human plasma defibrinated under controlled conditions, using

purified human thrombin. Plasma contains 20 mM HEPES.

Storage:

Plasma is shipped frozen and should be stored below-60°C.

Presentation:

Frozen Fibrinogen deficient plasma.

Specifications: Assay Results;

Activity Assay	Lot#DP1-0026
PT (HemosIL PT-Fib)	>120 sec
APTT (Siemens Actin FS)	>180 sec
FII (activity)	0.92 U/ml
FV (activity)	0.57 U/ml
FVIII (activity)	0.94 U/ml
Fibrinogen (antigen) by ELISA	0.05 g/L

Although this material was prepared with plasma collected from donors screened for CJD and which was tested at source and found negative for HBsAg, RPR and antibodies to HIV and HCV and non-reactive for HIV-1 rNA and HCV rNA by FDA approved tests, it should be handled by personnel trained in the proper procedures for handling potential viral contaminants.

For Research Purposes Only

10 mg vial
0.5 mg vial
0.2 mg vial
0.1 mg vial
10 mg vial
0.5 mg vial
0.2 mg vial
5 x 96 wells
4 x 96 wells
4 x 96 wells
1 mL to bulk
1 mL to bulk

Authorization:

Ada Blythe, Operations Manager

Date: Ola DIII

affinitybiologicals.com

Document OPI0338 rev. 3



