



Bio-X Diagnostics

BIO-X ABORTION ELISA KIT

Kit for serodiagnosis of BoHV-1 (IBR/IPV), BVDV, BHV-4 infections

Indirect test for blood sera

in cattle by ELISA

I – INTRODUCTION

The task of determining the cause of an abortion in cattle is generally a rather difficult one because, most of the time, it is the consequence of an event which happened weeks to months earlier. Often also, the foetus is maintained in the uterus for hours and even days after its death, and, when it is finally evacuated, it has undergone autolysis, in such a way that it is difficult to do any type of analysis. Also, many causes of abortion in cattle are to this day still unknown. Moreover, many pathogens are rarely looked for because they are difficult or dangerous to handle (*Coxiella burnetii*, *Chlamydia psittaci* ..). Pathogens directly or indirectly responsible for abortions are numerous and varied, which complicates the diagnosis. Amongst the major pathogens responsible for abortions, one can find viruses (BoHV-1, BVDV, BHV-4), bacteria (*Brucella abortus*, *Corynebacterium pyogenes*, *colibacille*, *streptocoque*, *Coxiella burnetii*, *Leptospira hardjo*, *Ureaplasma diversum*, *Campylobacter foetus*, *Borrelia coriaceae*, *Yersinia pseudotuberculosis*, *Chlamydia psittaci*, *Salmonella*, *Listeria monocytogenes* et *Haemophilus somnus*), parasites (*Distoma hepatica*, *Trichomonas*, *Sarcocystis*, *Neospora*), fungi (*Aspergillus fumigatus*, *Mortierella wolfii*, as well as *Mucor*, *Absidia*, *Rhizopus*) and yeasts (*Candida*).

The Bio-X Abortion ELISA kit concerns the three viruses BVDV, BHV-4 and BoHV-1. This kit aims at demonstrating the existence of a seroconversion toward the three viruses mentioned above, in adult cattle, that it is, in the animal who aborted, but especially in the other animals of the herd, ideally in 10% of the livestock or the cowshed. Indeed, when the abortion occurs, the serological titer of the cow has often reached its maximum and it is not possible to show a seroconversion. It is thus preferable to test the other animals of the herd in order to verify whether the suspected infection is still active. If many animals show a clear seroconversion toward one of the three pathogens of the kit, one can attribute the abortion to this pathogen.

II – PRINCIPLE OF THE TEST

The test uses 96-well microtitration plates sensitised by the three pathogens listed above. The distribution of these pathogens on the microtitration plate is as follows:

Rows A et E : BoHV-1,
Rows B et F : BVDV,
Rows C et G : BHV-4,
Rows D et H : negative control.

Rows D and H contain a lysate of the bovine kidney cell line that was used as a substrate to propagate the viruses. We thus have a genuine negative control to differentiate the virus-specific antibodies from those directed against the antigenic determinants of the kidney cells used for their replication. Using such a control reduces the number of false positives considerably. The test sera are diluted in an appropriate buffer and incubated on the plate for one hour at 21°C +/- 3°C. The plate is washed and the conjugate -a peroxidase-labelled anti-bovine IgG1 monoclonal antibody- is added to the wells. The plate is reincubated at 21°C +/- 3°C for 1 hour. After this second incubation, the preparation is washed and the chromogen (tetramethylbenzidine) is added. This chromogen has the advantages of being more sensitive than the other peroxidase chromogens and not being carcinogenic. If specific immunoglobulins are present in the test sera the conjugate remains bound to the corresponding microwell and the enzyme catalyses the transformation of the colourless chromogen into a pigmented compound. The intensity of the resulting blue colour is proportionate to the titre of specific antibody

in the sample. The signals recorded for the negative control microwells are subtracted from the corresponding positive microwells.

It is possible to quantify the reactivity of an unknown serum on a scale ranging from 0 to +++++.

III - COMPOSITION OF THE KIT

- **Microplates:** Two 96-well microtitration plates. The distribution of the different valencies on the microtitration plate is as follows :
Row A: BoHV-1
Row B: BVDV
Row C: BHV-4
Row D: negative control
Row E: BoHV-1
Row F: BVDV
Row G: BHV-4
Row H: negative control
- **Washing solution:** One bottle of 20x concentrated washing solution. The solution crystallises spontaneously when cold. If only part of the solution is to be used, bring the bottle to 21°C +/- 3°C so that all the crystals disappear. Mix the solution well and remove the necessary volume. Dilute the buffer 1:20 with distilled or demineralised water.
- **Dilution buffer:** One bottle of 5x concentrated buffer for diluting the blood sera and conjugate. The bottle's content is to be diluted with distilled or demineralised water. If a deposit forms at the bottom of the receptacle filter the solution on Whatman filter paper.
- **Conjugate:** One bottle of anti-bovine immunoglobulin-peroxidase conjugate (horseradish peroxidase-labelled anti-bovine IgG1 monoclonal antibody).
- **Positive serum:** One bottle of positive serum. Reconstitute this serum with 0.5 ml of distilled or demineralised water. The reconstituted serum may be kept at -20°C. Divide the reconstituted serum into several portions before freezing in order to avoid repeated freezing and thawing. If these precautions are taken the reagent may be kept for several months.
- **Single component TMB:** One bottle of the chromogen tetramethylbenzidine (TMB). Store between +2°C and +8°C protected from the light. Ready to use.
- **Stopping solution:** One bottle of the 1 M phosphoric acid stop solution.

	BIO K 072/2
Microplates	2
Washing solution	1 X 100 ml (20 X)
Dilution buffer	1 X 50 ml (5 X)
Conjugate	1 X 0,5 ml (50 X)
Positive serum	1 X 0,5 ml (1 X) freeze-dried
Single component TMB	1 X 25 ml (1 X)
Stop solution	1 X 15 ml (1 X)

IV - ADDITIONAL MATERIALS AND EQUIPMENT REQUIRED

Distilled water, graduated cylinders, beakers, plastic tubes, tube rack, dispenser tips, reagent reservoir for multichannel pipettes, lid, adhesive for microplates, graduated automatic (mono- and multichannel) pipettes, microplate reader, and microplate washer and shaker (optional).

V - PRECAUTIONS FOR USE

- This test may be used for “in vitro” diagnosis only. It is strictly for veterinary use.
- The reagents must be kept between +2°C and +8°C. The reagents cannot be guaranteed if the shelf-life dates have expired or if they have not been kept under the conditions described in this insert.
- The concentrated wash solution and dilution buffer may be stored at room temperature. Once diluted, these solutions remain stable for six weeks if kept between +2°C and +8°C.

- Unused strips must be stored immediately in the aluminium envelope, taking care to keep the desiccant dry and the envelope's seal airtight. If these precautions are taken, the strips' activity can be conserved up to the kit's shelf-life date.
- Do not use reagents from other kits.
- The quality of the water used to prepare the various solutions is of the utmost importance. Do not use water that may contain oxidants (e.g., sodium hypochlorite) or heavy metal salts, as these substances can react with the chromogen.
- Discard all solutions contaminated with bacteria or fungi.
- The stop solution contains 1 M phosphoric acid. Handle it carefully.
- All materials and disposable equipment that come in contact with the samples must be considered potentially infectious and be disposed of in compliance with the legislation in force in the country.
- To guarantee the reliability of the results, one must follow the protocol to the letter. Special care must be taken in observing the incubation times and temperatures, as well as measuring the volumes and dilutions accurately.

VI – PROCEDURE

- 1- Bring all reagents at 21°C +/- 3°C before use
- 2- Remove the microplate from its wrapper.
- 3- Transfer 1 ml/tube of the buffer for dilution, prepared as instructed in the "Composition of the Kit" section, to 5- or 10-ml hemolysis tubes. Add 10 µl of the serum samples to each of these tubes and shake briefly on a mechanical agitator. Proceed in the same manner for the positive serum.
- 4- Add 100-µl aliquots of the dilute samples (1:100) to the wells as follows: positive reference serum: wells A1 to D1, sample 1 wells E1 to H1, sample 2 wells A2 to D2, etc. Incubate the plate at 21°C +/- 3°C for one hour. Cover the plate with a lid.
- 5- Rinse the plate with the washing solution, prepared as instructed in the "Composition of the Kit" section, as follows: Empty the microtitration plate of its contents by flipping it over sharply above a sink. Tap the plate upside down against a piece of clean absorbent paper to remove all the liquid. Fill the utilised wells with the washing solution using a squeeze bottle or by plunging the plate in a vessel of the right dimensions, then empty the wells once more by turning the plate over above a sink. Repeat the entire operation two more times, taking care to avoid the formation of bubbles in the microwells. After the plate has been washed three times go on to the next step. An automatic plate washer may also be used, but in this case particular care must be taken to avoid any contact between the needles and the bottom of the wells to prevent any damage of the reagent layer.
- 6- Dilute the conjugate 1:50 in the buffer for dilution (for example, for one plate dilute 250 µl of the conjugate stock solution in 12.250 ml of diluent). Add 100 µl of the dilute conjugate solution to each well. Incubate for 1 hour at 21°C +/-3°C. Cover the plate with a lid.
- 7- Wash the plate as described in step 5 above.
- 8- Add 100 µl of the chromogen solution to each well on the plate. The chromogen solution must be absolutely colourless when it is pipetted into the wells. If a blue colour is visible, this means that the solution in the pipette has been contaminated.
- 9- Incubate for 10 minutes at 21°C +/-3°C. This time is given as a guideline only, for in some circumstances it may be useful to lengthen or shorten the incubation time.
- 10- Add 50 µl of stop solution per microwell.
- 11- Read the optical densities in the microwells using a plate reader and a 450 nm filter. Results must be read fairly soon after the stopping solution has been added since the chromogen may crystallize in wells with strong signals and thereby distort the data.

VII – INTERPRETING THE RESULTS

Subtract from each value recorded in rows A, B, C the signal of the corresponding negative control (row D) and for the rows E, F, G the signal from the row H. Write down the result. Allow for any negative values that may exist in performing this calculation Carry out the same operations for the column corresponding to the positive control (wells A1 to D1).

The test can be validated only if the positive serum yields a difference in optical density at 10 minutes that is greater for each valence than :

BoHV-1: 1,000
BVDV : 0,600
BHV-4 : 0,800

Divide the signal read for each sample well by the corresponding positive control serum signal and multiply this result by 100 to express it as a percentage.

$$\text{Val} = \frac{\text{Delta DO éch} * 100}{\text{Delta DO pos}}$$

Using the following table, determine each serum's degree of positivity.

	0		+		++		+++		++++		+++++
BoHV-1	Val <=	20 %	< Val <=	45 %	< Val <=	60 %	< Val <=	85 %	< Val <=	110 %	< Val
BVDV	Val <=	37 %	< Val <=	74 %	< Val <=	111 %	< Val <=	148 %	< Val <=	185 %	< Val
BHV-4	Val <=	27 %	< Val <=	54 %	< Val <=	81 %	< Val <=	108 %	< Val <=	135 %	< Val

A reliable diagnosis can be made only if frank seroconversion can be documented using two coupled serum samples taken at 2- to 3-week intervals. The first sample must be taken during the acute phase of the infection. A frank seroconversion is considered to have occurred if the signal increases by two orders of magnitude (two plusses; for example, ++ -> ++++ or + -> +++). A sample must be considered positive if it yields a result that is greater than or equal to one plus sign (+).

VII – ORDERING INFORMATION

BIO-X ABORTION ELISA KIT :

3 x 48 tests

BIO K 072/2

