

Polink DS-MM D kit for for Immunohistochemistry Staining

**Polymer-HRP&AP double staining kit to distinct two mouse antibodies
on Human tissue with DAB (Brown) and Fast Red (Red)**

Storage: 2-8°C

Catalog No.: ☐ DS203D-6/(D78-6F) 12ml* 60 slides**

☐ DS203D-18 36ml* 180 slides**

☐ DS203D-60 120ml* 600slides**

*Total volume of polymer Conjugates

** if use 100µl per slide

Intended Use:

The **Polink DS-MM D Kit** is designed to use with user supplied two mouse antibodies to detect two distinct antigens on human tissue or cell samples. This kit has been tested in paraffin tissue. However, this kit can be used on frozen specimen and freshly prepared monolayer cell smears.

Double staining is one of most common methods used in immunohistostaining that allow revealing two distinct antigens in a single tissue^{1, 2}. **Polink DS-MM Kit** from GBI Labs (Golden Bridge International) supplies two polymer enzyme conjugates: HRP polymer anti-Mouse IgG and AP polymer anti-Mouse IgG with two distinct substrates/chromogens, DAB (brown color, use with HRP polymer anti-Mouse IgG) and Fast Red (red color, use with AP polymer anti-Mouse IgG). **Polink DS-MM Kit** is non-biotin system that avoids endogenous biotin non-specific binding.

Kit Components:

Component No.	Content	12ml Kit	36ml Kit	120ml Kit
Reagent 1	HRP polymer anti-Mouse IgG (RTU)	6ml	18ml	60ml
Reagent 2A	DAB substrate buffer (RTU)	6ml	18ml	60ml
Reagent 2B	DAB chromogen (20x)	1ml	1ml	3ml
Reagent 3A	DS-MM Blocker A (RTU)	6ml	18ml	60ml
Reagent 3B	DS-MM Blocker B (RTU)	6ml	18ml	60ml
Reagent 4	AP polymer anti-Mouse IgG (RTU)	6ml	18ml	60ml
Reagent 5A	Fast Red chromogen tablets	6 tablets	18 tablets	60 tablets
Reagent 5B	Fast Red substrate buffer	5ml x 6	5ml x 18	5ml x 60
Reagent 6	Simplo-Mount solution (RTU)	6ml	18ml	60ml

Recommended Protocol:

1. Fixation: To ensure the quality of the staining and obtain reproducible performance, user needs to supply appropriately fixed tissue and well prepared slides.
2. Tissue need to be adhered to the slide tightly to avoid tissue falling off.
3. Paraffin embedded section must be deparaffinized with xylene and rehydrated with a graded series of ethanol before staining.
4. Cell smear samples should be made as much monolayer as possible to obtain satisfactory results.
5. Three control slides will aid the interpretation of the result: positive tissue control, reagent control (slides treated with Isotype control reagent), and negative control.
6. It takes about 30 minutes to dissolve Fast Red tablet into the substrate buffer. Make sure to start preparing Fast Red solution near the end of the secondary antibody incubation.
7. Proceed IHC staining: DO NOT let specimen or tissue dry from this point on.

Reagent	Staining Procedure	Incubation Time (Min.)
1. Peroxidase Blocking Reagent Not provided	a. Incubate slides in peroxidase blocking reagent (Ready-to-use 3% H ₂ O ₂ solution) for 10 minutes. b. Rinse the slide using distilled water.	10 min
2. HIER Pretreatment: Refer to antibody data sheet.	c. Heat Induced Epitope Retrieval (HIER) may be required for primary antibody suggested by vendor. d. Wash with PBS for 2 min., 3 times.	
3. Preblock	For paraffin section, Improved formula saves the need for a preblock	

(optional)	step. For frozen tissue, preblock may or may not be required depending on fixative. (Preblock catalogue No.:E07 was Recommended.)	
4. Mouse Antibody 1: Supplied by user	Notes: Investigator needs to optimize dilution and incubation times prior to double staining. a. Apply 2 drops or enough volume of mouse primary antibody 1 to cover the tissue completely. Incubate in moist chamber for 30-60 min. b. Rinse with PBS containing 0.05% Tween-20 for 2 min.,3 times.	30-60 min
5. Reagent 1: HRP polymer anti-Mouse IgG(RTU)	a. Apply 1drop (50ul) of Reagent 1 HRP polymer anti-Mouse IgG to cover each section. b. Incubate in moist chamber for 15-30 min. c. Rinse with PBS containing 0.05% Tween-20 for 2 min.,3 times.	15-30 min
6. Reagents 2A, 2B: Reagents 2A: DAB Substrate Reagents 2B: DAB chromogen (20x)	a. Add 1 drop or 2 drops (for higher sensitivity and contrast) of Reagent 2B to 1 ml Reagent 2A . Mix well. Protect from light and use within 5 hours. b. Apply 2 drops or enough volume of DAB CHROMOGEN to completely cover tissue. Incubate for 3-10 min. c. Rinse thoroughly with distilled water 4 times, 2 minutes each time.	3-10 min
7. Reagent 3A: DS-MM Blocker A	a. Apply 2 drops or enough volume of Reagent 3A DS-MM Blocker A to cover the tissue completely. Mix well on the slide and Incubate in moist chamber for 30 min. b. Rinse with PBS containing 0.05% Tween-20 for 2 min.,3 times.	30 min
7. Reagent 3B: DS-MM Blocker B	c. Apply 2 drops or enough volume of Reagent 3B DS-MM Blocker B to cover the tissue completely. Mix well on the slide and Incubate in moist chamber for 5 min. d. Rinse with PBS containing 0.05% Tween-20 for 2 min.,3 times.	5 min
8. Mouse antibody 2: Supplied by user	Notes: Investigator needs to optimize dilution and incubation times prior to double staining. a. Apply 2 drops or enough volume of mouse primary antibody 2 to cover the tissue completely. b. Rinse with PBS containing 0.05% Tween-20 for 2 min.,3 times.	30-60 min
9. Reagent 4: AP polymer anti-Mouse IgG(RTU)	a. Apply 1drop (50ul) of Reagent 4 AP polymer anti-Mouse IgG to cover each section. b. Incubate in moist chamber for 15-30 min. c. Rinse with PBS containing 0.05% Tween-20 for 2 min.,3 times. d. Rinse well with tap water.	15-30 min
10. Reagent 5A, 5B: Reagent 5A: Fast Red chromogen tablets Reagent 5B: Fast Red substrate buffer	Notes: It takes about 30 minutes to dissolve the tablet in the substrate buffer. Allow enough time to prepare. a. Dissolve 1 Reagent 5A Fast Red tablet in 5ml Reagent 5B Fast Red substrate buffer, vortex until the tablet dissolved completely. Use within 1 hour. b. Apply 2 drops (100 µl) or enough volume of Fast -Red solution to completely cover the tissue. Incubate for 10-20 min, observe appropriate color development c. Rinse well with distilled water. (Fast Red is alcohol soluble; do not dehydrate.)	10-20 min
11. HEMATOXYLIN Not provided	a. Counterstain with 2 drops (100 µl) or enough volume of hematoxylin to completely cover tissue. Incubate for 10-15 seconds. b. Rinse thoroughly with tap water for 2-3 min c. Put slides in PBS until show blue color (about ½ - 1 min.) d. Rinse well in distilled water	
12. Reagent 6: Simpomount	a. Apply 2 drops (100 µl) or enough volume to cover tissue when tissue is wet. Rotate the slides to allow Simpo-Mount spread evenly. DO NOT coverslip. b. Place slides horizontally in an oven at 40-50°C for at least 30 minutes or leave it at room temperature until slides are thoroughly dried. Hardened Simpo-Mount forms an impervious polymer barrier to organic solvent. Do not use oil directly on	30 min. in 40-50°C oven Or: overnight at room temperature

	the top of dried Simpo-Mount.	
--	-------------------------------	--

Protocol Notes:

1. The fixation, tissue slide thickness, antigen retrieval and primary antibody dilution and incubation time effect results significantly. Investigator needs to consider all factors and determine optimal conditions when interpret the result.
2. Simpo-Mount is water-based mounting medium for immunohistology. It may be used as the permanent mounting media. It does not need coverslip. However, if you need to coverslip, after Simpo-Mount dried, dip the slide in xylene and take out immediately. Apply O-Mount (organic mounting solution, Catalog No. E02-15) on the tissue and place cover glass on the slide. Store it after dry completely.

Precautions:

DAB may be carcinogenic. Please wear gloves and take other necessary precautions.

Remarks:

For research use only.

References:

1. De Pasquale A, Paterlini P, Quaglino D. Immunochemical demonstration of different antigens in single cells in paraffin-embedded histological sections. Clin Lab Haematol. 1982;4(3):267-72.
2. Polak J. M and Van Noorden S. Introduction to Immunocytochemistry Second Edition. Bios Scientific Publishers. P41-54. 1997