

MOS #NBTH-100 , 500 , 1000

MOSS, INC.  
P.O. BOX 189  
PASADENA, MD 21123  
1-800-932-6677  
(410)-768-3442  
FAX (410)-768-3971

\*\*\*CERTIFICATE OF ANALYSIS\*\*\*

PRODUCT #NBTH-100, -500, -1000

BCIP/NBT (HISTO)  
5-bromo-4-chloro-3-indolyl-phosphate  
p-nitro blue tetrazolium chloride in  
a proprietary buffer and enhancer

EXPIRATION DATE: JAN 2006

LOT# 07189041

COLOR & CLARITY: --clear to pale yellow     \_X\_pass     \_\_\_fail

pH 9.8 +/- 0.1                             \_X\_pass     \_\_\_fail

STABILITY @ 37 °C :                     \_X\_pass     \_\_\_fail

STABILITY @ 4 °C :                     \_X\_pass     \_\_\_fail

STABILITY @ RT :                         \_X\_pass     \_\_\_fail

Spectral scan of raw materials where required : \_X\_pass     \_\_\_fail

Spectral scan of finished product :     \_X\_pass     \_\_\_fail

Performance data on "dot blot" :       \_X\_pass     \_\_\_fail

Check 18 mega Ohm water monitor prior to production: \_X\_

Sterile filtered water .22u :           \_X\_

Storage conditions: Room temperature or Refrigerate.

Shipping Conditions : regular - no special precautions necessary.

RON TELFORD  
Mgr QA/QC

コスモ・バイオ株式会社