

MOSS, INC.
P.O. BOX 189
PASADENA, MD 21123
1-800-932-6677
(410)-768-3442
FAX (410)-768-3971

CERTIFICATE OF ANALYSIS

PRODUCT #NBTM-100, -500, -1000, -50BTL, -25BTL

BCIP/NBT
5-bromo-4-chloro-3-indolyl-phosphate
p-nitro blue tetrazolium chloride in
a proprietary buffer and enhancer

EXPIRATION DATE: NOV 2005
MAY 2006 (if stored at 2°C - 8°C upon arrival)

LOT# 05140041

COLOR & CLARITY: -clear to pale yellow _X_pass _fail

pH 9.8 +/- 0.1 _X_pass _fail

STABILITY @ 4 °C : _X_pass _fail

STABILITY @ RT : _X_pass _fail

Spectral scan of raw materials where required : _X_pass _fail

Performance data on "dot blot" : _X_pass _fail

Check 18 mega Ohm water monitor prior to production: _X_

Sterile filtered water .22u : _X_
Storage conditions: 18°C - 26°C or REFRIGERATE 2°C - 8°C
Shipping Conditions : regular - no special precautions necessary.

Ron Telford

Manager QC/QA

コスモ・バイオ株式会社