



Telephone: 509-335-6058
Fax 509-335-8328
e-mail mabcenterorders@vetmed.wsu.edu
web http://www.vetmed.wsu.edu/depts-MoAb/



## **Certificate of Analysis**

Catalog No. PG2005/cell line: TH22A


Specificity: MHC II Isotype: IgG2a

Known applications: Flow cytometry, immunohistochemistry, cell function

Known species reactivity: Pig, bovine, caprine, ovine

Description: The monoclonal antibody was produced in ascites fluid, clarified by filtration through a 0.2 micrometer filter. The antibody concentration is 1.0 mg/ml in phosphate buffered saline, preserved in 0.09% sodium azide. Pig, bovine, caprie, ovine

Method used for quality control: Flow cytometry



Flow cytometric profile obtained with pig leukocytes

**STORAGE:** When the vial is stored at 2-7°C, it should be stable for one year.

## **REFERENCES:**

W. C. Davis, S. Marusic, H. A. Lewin, G. A. Splitter, L. E. Perryman, T. C. McGuire, and J. R. Gorham. The development and analysis of species specific and cross reactive monoclonal antibodies to leukocyte differentiation antigens and antigens of the major histocompatibility complex for use in the study of the immune system in cattle and other species. *Vet.Immunol.Immunopathol.* 15:337-376, 1987.

A. Ababou, J. Goyeneche, W. C. Davis, and D. Levy. Evidence for the expression of three different BoLa-class II molecules on the bovine BL-3 cell line: determination of a non-DR non-DQ gene product. *J.Leukoc.Biol.* 56:182-186, 1994.

## For *in vitro* research use only

Warranty: The WSU Monoclonal Antibody Center warrants this monoclonal antibody is as described and present in the amount stated on the label at the time of delivery to the customer. No other warranties, expressed or implied are made beyond the label description, including warranties of merchantability or fitness for a particular use. Remedy is limited to replacement of the product. The Center is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in the data sheet is provided for reference only and not a substituted for the user's own respective material quality control.